



No. ST/NARFBR/EEG Machine/NHP/P49/2024-25

Date: 22.01.2026

Reference is invited to the Tender Notice / Bid through GeM Portal *vide* Bid No: GEM/2025/B/7045725 dt.01.01.2026 End dt: 22.01.2026.

As per the tender/ Bid document (Technical Specification)	Revised specification To be read as
Sl. No. 5 (iii) under EEG Software features Automatic Artifact detection (eye blink, ECG, EMG) preferred	Automatic/manual artifact detection (eye blink, ECG, EMG) preferred.
Sl. No. 3 Under EMG unit EMG amplifier should have control knobs for: Gain, sensitivity, cursors, pulse width and stimulus parameters	EMG amplifier/software should have control for: Gain, sensitivity, cursors, pulse width and stimulus parameters.

**Note: Remaining all Specifications & terms & conditions will remain unchanged.**

Administrative Officer